

SSF BAT2 Microphone exchange

On SSF-BAT2 devices, where the serial number ends with "S" or is marked with a green dot (the serial number is located in the battery compartment), the microphone can easily be replaced by the user via a connector system. Older devices must be sent to the company microelectronic Volkmann for the microphone replacement.

Following tools are needed:


1 small Phillips screwdriver for the removing of the housing screws

1 small screwdriver as a lever tool

Remove the batteries before the conversion!




Undo the 4 housing screws



Carefully lifting the circuit board,
unplug the microphone to the front.



Ultrasonic sensor



Place the ultrasonic sensor on the plug,
white marking points upwards!



Press the circuit board down,
close the housing and screw it tight.
Ready for Usage...

CAUTION:

Please pay attention to the sensitive electronics when opening the device and working on it. The conversion takes place on your own risk. No guarantee can be given.

If you are not confident in doing this microphone exchange, please send the device to the company for the exchange to be done.